



UNITED STATES PATENT AND TRADEMARK OFFICE

me
UNITED STATES DEPARTMENT OF COMMERCE
United States Patent and Trademark Office
Address: COMMISSIONER FOR PATENTS
P.O. Box 1450
Alexandria, Virginia 22313-1450
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/023,405	12/14/2001	Stephen C. Wardlaw	5169-04-1-1	8490

7590 02/11/2004

RICHARD D. GETZ, Esq.
McCormick, Paulding & Huber LLP
City Place II
185 Asylum Street
Hartford, CT 06103-3402

EXAMINER

SINES, BRIAN J

ART UNIT	PAPER NUMBER
----------	--------------

1743

DATE MAILED: 02/11/2004

Please find below and/or attached an Office communication concerning this application or proceeding.

Office Action Summary	Application No.	Applicant(s)	
	10/023,405	WARDLAW, STEPHEN C.	
	Examiner	Art Unit	
	Brian J. Sines	1743	

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --

Period for Reply

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If the period for reply specified above is less than thirty (30) days, a reply within the statutory minimum of thirty (30) days will be considered timely.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133).
- Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

Status

- 1) ☒ Responsive to communication(s) filed on 22 December 2003.
- 2a) ☐ This action is **FINAL**. 2b) ☒ This action is non-final.
- 3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

Disposition of Claims

- 4) ☒ Claim(s) 1-9 is/are pending in the application.
- 4a) Of the above claim(s) 8 and 9 is/are withdrawn from consideration.
- 5) ☐ Claim(s) _____ is/are allowed.
- 6) ☒ Claim(s) 1-7 is/are rejected.
- 7) ☐ Claim(s) _____ is/are objected to.
- 8) ☐ Claim(s) _____ are subject to restriction and/or election requirement.

Application Papers

- 9) ☐ The specification is objected to by the Examiner.
- 10) ☐ The drawing(s) filed on _____ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.
 Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).
 Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).
- 11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

Priority under 35 U.S.C. §§ 119 and 120

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).
 a) ☐ All b) ☐ Some * c) ☐ None of:
 1. ☐ Certified copies of the priority documents have been received.
 2. ☐ Certified copies of the priority documents have been received in Application No. _____.
 3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).
 * See the attached detailed Office action for a list of the certified copies not received.
- 13) ☒ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. § 119(e) (to a provisional application) since a specific reference was included in the first sentence of the specification or in an Application Data Sheet. 37 CFR 1.78.
 a) ☐ The translation of the foreign language provisional application has been received.
- 14) ☐ Acknowledgment is made of a claim for domestic priority under 35 U.S.C. §§ 120 and/or 121 since a specific reference was included in the first sentence of the specification or in an Application Data Sheet. 37 CFR 1.78.

Attachment(s)

- | | |
|---|---|
| 1) <input checked="" type="checkbox"/> Notice of References Cited (PTO-892) | 4) <input type="checkbox"/> Interview Summary (PTO-413) Paper No(s). _____ |
| 2) <input type="checkbox"/> Notice of Draftsperson's Patent Drawing Review (PTO-948) | 5) <input type="checkbox"/> Notice of Informal Patent Application (PTO-152) |
| 3) <input checked="" type="checkbox"/> Information Disclosure Statement(s) (PTO-1449) Paper No(s) _____ | 6) <input type="checkbox"/> Other: _____ |

Art Unit: 1743

DETAILED ACTION***Election/Restrictions***

Applicant's election of group I, claims 1 – 7 in the response received 12/22/2003 is acknowledged. Because applicant did not distinctly and specifically point out the supposed errors in the restriction requirement, the election has been treated as an election without traverse (MPEP § 818.03(a)). Claims 8 and 9 are withdrawn from further consideration pursuant to 37 CFR 1.142(b) as being drawn to a nonelected invention, there being no allowable generic or linking claim.

Claim Rejections - 35 USC § 112

The following is a quotation of the second paragraph of 35 U.S.C. 112:

The specification shall conclude with one or more claims particularly pointing out and distinctly claiming the subject matter which the applicant regards as his invention.

Claims 1 – 7 are rejected under 35 U.S.C. 112, second paragraph, as being indefinite for failing to particularly point out and distinctly claim the subject matter which applicant regards as the invention. Regarding claims 1, 3 and 7, it is unclear as to what specifically are the operable features which enable the testing of the biological fluid samples. The applicant is advised that although the claims are interpreted in light of the specification, limitations from the specification are not read into the claims. See *In re Van Geuns*, 988 F.2d 1181, 26 USPQ2d 1057 (Fed. Cir. 1993). “The PTO applies to the verbiage of the proposed claims the broadest reasonable meaning of the words in their ordinary usage as they would be understood by one of ordinary skill in the art.” See *In re Morris*, 127 F.3d 1048, 1054, 44 USPQ2d 1023, 1027 (Fed. Cir. 1997). “During patent examination the pending claims must be interpreted as broadly as their terms reasonably allow.” See *In re Zletz*, 893 F.2d 319, 321-22, 13 USPQ2d 1320, 1322 (Fed. Cir.

Art Unit: 1743

1989). "The PTO broadly interprets claims during examination of a patent application since the applicant may 'amend his claim to obtain protection commensurate with his actual contribution to the art.'"(quoting *In re Prater*, 415 F.2d 1393, 1404-05, 162 USPQ 541, 550 (CCPA 1969)). See *In re Yamamoto*, 740 F.2d 1569, 1571, 222 USPQ 934, 936 (Fed. Cir. 1984). The Courts have held that apparatus claims must be structurally distinguishable from the prior art in terms of structure, not function. See *In re Danley*, 120 USPQ 528, 531 (CCPA 1959); and *Hewlett-Packard Co. V. Bausch and Lomb, Inc.*, 15 USPQ2d 1525, 1528 (Fed. Cir. 1990). The Courts have held that the manner of operating an apparatus does not differentiate an apparatus claim from the prior art, if the prior art apparatus teaches all of the structural limitations of the claim. See *Ex Parte Masham*, 2 USPQ2d 1647 (BPAI 1987) (see MPEP § 2114). Therefore, with regards to the specification of the particular features operable to enable the testing of the biological fluid, the term "feature" is being broadly and reasonably interpreted in the art for examination purposes as being a feature, such as a test well as taught by Merkh *et al.*, which enables the testing of the fluid sample. Although the apparatus as taught by the prior art may not be what the applicant intends as their claimed invention, the claims still encompass the teachings of the prior art. In other words, the claims do not *exclude* the teachings of the prior art. Furthermore, regarding claim 5, these operable features as shown by the specification, appear to be critical to the functioning of the apparatus. The applicant is advised that a feature, which is taught as critical in the specification, should be recited in the claims (see MPEP § 2164.08 (c)). The structure which goes to make up the device must be clearly and positively specified. The structure must be correlated in such a manner as to present a complete operative device.

Art Unit: 1743

Claim Rejections - 35 USC § 102

The following is a quotation of the appropriate paragraphs of 35 U.S.C. 102 that form the basis for the rejections under this section made in this Office action:

A person shall be entitled to a patent unless –

(b) the invention was patented or described in a printed publication in this or a foreign country or in public use or on sale in this country, more than one year prior to the date of application for patent in the United States.

Claims 1 and 3 are rejected under 35 U.S.C. 102(b) as being anticipated by Merkh *et al.* (U.S. Pat. No. 5,281,540 A). Merkh *et al.* teach an apparatus comprising: a chamber (86) for holding a sample; features (*e.g.*, wells 84) operable to enable the testing of the sample; a label (94) containing information for the performance of tests; a label reader (32); a field illuminator (optical source means 162); an image dissector (optical detector means 164); and a positioner (*e.g.*, reader arm 35 or carousel 18) (see col. 8, lines 7 – 37 & col. 21, lines 20 – 63).

Allowable Subject Matter

Claims 2 and 4 – 7 would be allowable if rewritten to overcome the rejection(s) under 35 U.S.C. 112, second paragraph, set forth in this Office action and to include all of the limitations of the base claim and any intervening claims.

The following is a statement of reasons for the indication of allowable subject matter: The cited prior art neither teach or fairly suggest the further incorporation of a means for determining either one of a through-plane thickness or volume of the sample field within the apparatus of Merkh *et al.* in order to facilitate sample testing.

Conclusion

The prior art made of record and not relied upon is considered pertinent to applicant's disclosure. Oberhardt teaches cell analysis methods, which rely on photodetection. Bacus *et al.*


Art Unit: 1743

teach methods and apparatus, which incorporate the use of image analysis in analyzing cell samples. Allen *et al.* teach an electronic assay device and method, which incorporate the use of an optical detection system.

Any inquiry concerning this communication or earlier communications from the examiner should be directed to Brian J. Sines, Ph.D. whose telephone number is (571) 272-1263. The examiner can normally be reached on Monday - Friday (11:30 AM - 8 PM EST).

If attempts to reach the examiner by telephone are unsuccessful, the examiner's supervisor, Jill A. Warden can be reached on (571) 272-1267. The fax phone number for the organization where this application or proceeding is assigned is (703) 872-9306.

Any inquiry of a general nature or relating to the status of this application or proceeding should be directed to the receptionist whose telephone number is not available.


Jill Warden
Supervisory Patent Examiner
Technology Center 1700